



Centers for Disease Control and Prevention (CDC)  
National Institute for Occupational Safety and Health (NIOSH)  
National Personal Protective Technology Laboratory (NPPTL)  
626 Cochrans Mill Road  
Pittsburgh, PA 15236-0070

January 10, 2023

NIOSH Reference: TN-26168  
AAR Number: LWS113

Peter Maguire  
Vice President  
Lighthouse Worldwide Solutions  
300 W Antelope Road  
White City, OR 97503

Dear Peter Maguire:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted November 9, 2022. This request was for approval of the SHIELD-95-PRO air-purifying filtering facepiece respirator for protections against particulates at a N95 filter efficiency level. The complete respirator configuration is detailed on assembly matrix, file name *714252843-1AM23.xls*, revision 23, dated: January 3, 2023.

This request is granted to 1 new approval and 3 modifications to existing approvals. Approvals are granted only for documentation written in the English language. It is the approval holder's responsibility to correctly translate materials desired to be produced in languages other than English. Approval number **TC-84A-9466** has been assigned. This respirator is approved for protection against particulates at a N95 filter efficiency level.

The approval label is included as an attachment to this letter. The abbreviated label has been accepted as submitted. The full and abbreviated approval labels must be presented to users as approved by NIOSH, including the cautions and limitations as listed. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assemblies consist of the parts as listed on the approval labels and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

Production and labeling of the approved respirator configuration can begin following receipt of this letter (the time and date of the approval action are recorded by my electronic signature, below).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely,

Jeffrey Peterson  
Chief, Conformity Verification and  
Standards Development Branch

Enclosures